





Klicka på "Download Data" för att ladda över mätningar till Pc'n





Välj "Megohmmeter" och "CA6549"

The dialog box titled "Select Instrument" contains two lists. The first list, labeled "Instrument", includes: Ground Tester, Installation Tester, Megohmmeter, Micro Ohmmeter, Power Analyzer, and Simple Logger II. The second list, labeled "Model", includes: C.A. 6543 - Megohmmeter 1000V, C.A. 6547 - Megohmmeter 5000V, and C.A. 6549 - Graphical Megohmmeter 5000V. The "Megohmmeter" item in the first list and the "C.A. 6549 - Graphical Megohmmeter 5000V" item in the second list are highlighted with red circles. "OK" and "Cancel" buttons are located to the right of the lists.



Välj rätt COM-port, som instrumentet är kopplad till

Connection ✕

Communications Port	COM3	OK
Communication Rate	9600	Cancel
		Help



Välj det Object och Test som skall laddas över till PC'n,

Select Tests

Select tests to download

- Object 1, Test 1
- Object 1, Test 2
- Object 2, Test 1
- Object 2, Test 2
- Object 3, Test 1
- Object 4, Test 1
- Object 4, Test 2
- Object 5, Test 1
- Object 6, Test 1
- Object 6, Test 2
- Object 7, Test 1
- Object 8, Test 1

Download

Cancel

Select All

Deselect All

Help



Specify (or accept) the name for Object 8, Tests 1



Välj "Manteltestrapport 2", för att generera en Rapport.

Create Report from Template [?] [X]

Step 1: Specify Template

Select Instrument	Select Template
GroundTester	Continuity Test Report
ICT	Manteltestrapport
Megohmmeter	ManteltestRapport_2
MicroOhmmeter	Megohmmeter CA6542 Summary Report
Qualistar	Megohmmeter CA6547 Summary Report
Simple Logger II	Megohmmeter CA6549 PI DAR
Wattmeter	Megohmmeter CA6549 Summary Report


or

C:\.....\Megohmmeter\ManteltestRapport_2.dvt [Browse]

Step 2: Specify Database(s)

1: C:\.....\Megohmmeter\Objekt8.dvb [Browse]

[OK] [Cancel] [Help]



Manteltest Rapport

2010-09-20

Operatör
CA Mätssystem AB
Mustafa Alpturk
Sjöflygvägen 35

Täby, Stockholm
070-769 11 94

Test Plats
CA Mätssystem AB
Mustafa Alpturk
Sjöflygvägen 35

Täby, Stockholm
070-769 11 94

Instrument ID: Model: CA6549, ID: 700652, Firmware: V20

Kommentarer
Testmätning utförd i demosyfte.

Database:
Object8.dwb

